

**Environmental Bio-Systems, Inc. Field Laboratory  
Spectrophotometry Analysis DR-2800 Methods  
Parameter/Range/Reagent Information**

NOTE: To view a test procedure, select the link in the "Proc" column.

**Alachlor**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Immunoassay	Threshold, low ppb		<a href="#">View</a>		

**Aluminum**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Chromazurol, Method 10215	0.02-0.50 mg/L Al		<a href="#">View</a>		<a href="#">Aluminum TNTplus (0.02-0.50 mg/L), 24 Tests</a>
Aluminon, Method 8012	0.008 to 0.800 mg/L		<a href="#">View</a>		<a href="#">Aluminum Reagent Set, AluVer® 3, 100/test</a> <a href="#">Aluminum Reagent, AluVer® 3, pk/100 Powder Pillows</a> , <a href="#">Ascorbic Acid, pk/100, Powder Pillows</a> , <a href="#">Bleaching 3 Reagent, pk/100 Powder Pillows</a>
Eriochrome Cyanine R, Method 8326	0.002 to 0.250 mg/L Al <sup>3+</sup>		<a href="#">View</a>		<a href="#">Aluminum Reagent Set, Eriochrome Cyanine R (ECR) Method, 10 mL sample, 100 tests</a> , <a href="#">ECR REAGENT PWD PLW 20ML PK/100</a> , <a href="#">HEXAMETHYLENETETRAMINE PK/100</a>

**Arsenic**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Silver Diethyldithiocarbamate, Method 8013	0-0.200 MG/l As		<a href="#">View</a>	USEPA	<a href="#">Arsenic Standard Solution as As, 1000 mg/L, Bottle/100 mL</a> , <a href="#">Hydrochloric Acid 500 mL</a> , <a href="#">Lead Acetate Solution 10% 100 mL</a> , <a href="#">POTASSIUM IODIDE SOLN, 20% 100ML</a> , <a href="#">Pyridine 500 mL</a> , <a href="#">Silver Diethyldithiocarbamate 25 g</a>

**Atrazine**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Immunoassay, Method 10050	Threshold, low ppb		<a href="#">View</a>		

**Barium**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Turbidimetric, Method 8014	2 to 100 mg/L		<a href="#">View</a>		<a href="#">Barium Reagent, BariVer® 4, pk/100 Powder Pillows</a>

**Boron**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Azomethine-H, Method 10061	0.02 to 1.50 mg/L as B		<a href="#">View</a>		<a href="#">Boron Reagent Set, BoroTrace Method, 0 to 1.50 mg/L boron, 100 tests</a> , <a href="#">BoroTrace(TM) 2 Reagent, pk/100 Powder Pillows</a> , <a href="#">EDTA SOLN, 1M 50ML SCDB</a>

**Bromine**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
DPD, Method 8016	0.05 to 4.50 mg/L		<a href="#">View</a>		<a href="#">DPD Total Chlorine Reagent, 10 mL Sample, Powder Pillows, PK/100</a> , <a href="#">DPD Total Chlorine Reagent, AccuVac® Ampules, PK/25</a>

**Cadmium**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Dithizone, Method 8017	0 to 80.0 µg/L		<a href="#">View</a>		<a href="#">Cadmium Reagent Set, 100 tests (No Air Shipments)</a>
Cadion, Method 10217	0.02-0.30 mg/L Cd		<a href="#">View</a>		<a href="#">Cadmium TNTplus (0.02-0.30 mg/L), 25 tests</a>

ENVIRONMENTAL BIO-SYSTEMS ([www.ebsinfo.com](http://www.ebsinfo.com); [jimjacobs@ebsinfo.com](mailto:jimjacobs@ebsinfo.com))

**Environmental Bio-Systems, Inc. Field Laboratory  
Spectrophotometry Analysis DR-2800 Methods  
Parameter/Range/Reagent Information**

**Chemical Oxygen Demand (COD)**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Manganese III Digestion, Method 10067	30 to 1000 mg/L COD Mn		<a href="#">View</a>		<a href="#">COD Digestion Vials, Manganese III, 20-1000 mg/L, pk/25</a> , <a href="#">Manganese III Chloride Removal Cartridges, pk/25</a> , <a href="#">Sulfuric Acid, ACS Grade, 4 kg</a>
Manganese III Reactor Digestion, Method 10067	30 to 1000 mg/L COD Mn		<a href="#">View</a>		<a href="#">COD Digestion Vials, Manganese III, 20-1000 mg/L, pk/25</a>
Reactor Digestion, Method 8000	3-150 mg/L COD		<a href="#">View</a>	USEPA , USEPA	<a href="#">TNTplus Chemical Oxygen Demand (COD) Reagent, Low Range (3-150 mg/L), 150 tests</a> , <a href="#">TNTplus Chemical Oxygen Demand (COD) Reagent, Low Range (3-150 mg/L), 25 tests</a> , <a href="#">COD Digestion Vials, Low Range, 3 to 150 mg/L COD, pk/25</a> , <a href="#">COD Digestion Vials, Low Range, 3 to 150 mg/L COD, pk/150</a>
Reactor Digestion, Method 8000	20-1500 mg/L COD		<a href="#">View</a>	USEPA , USEPA	<a href="#">TNTplus Chemical Oxygen Demand (COD) Reagent, High Range (20-1500 mg/L), 150 tests</a> , <a href="#">TNTplus Chemical Oxygen Demand (COD) Reagent, High Range (20-1500 mg/L), 25 tests</a> , <a href="#">COD Digestion Vials, High Range, 20 to 1,500 mg/L COD, pk/150</a> , <a href="#">COD Digestion Vials, High Range, 20 to 1,500 mg/L COD, pk/25</a>
Reactor Digestion, Method 8000	200 - 15,000 mg/L COD				<a href="#">COD Digestion Vials, High Range Plus, 200 to 15,000 mg/L COD, pk/150</a> , <a href="#">COD Digestion Vials, High Range Plus, 200 to 15,000 mg/L COD, pk/25</a>
Reactor Digestion, Method 8000	3 - 150 mg/L COD				<a href="#">COD Digestion Vials, Ultra Low Range, 1 to 40 mg/L COD, pk/150</a> , <a href="#">COD Digestion Vials, Ultra Low Range, 1 to 40 mg/L COD, pk/25</a>
COD, Method 10211	1-60 mg/L		<a href="#">View</a>		<a href="#">TNTplus Chemical Oxygen Demand (COD) Reagent, Ultra Low Range (1-60 mg/L), 144 tests</a> , <a href="#">TNTplus Chemical Oxygen Demand (COD) Reagent, Ultra Low Range (1-60 mg/L), 24 tests</a>
COD, Method 10212	250-15000 mg/L		<a href="#">View</a>		<a href="#">TNTplus Chemical Oxygen Demand (COD) Reagent, Ultra High Range (200-15,000 mg/L), 150 tests</a> , <a href="#">TNTplus Chemical Oxygen Demand (COD) Reagent, Ultra High Range (200-15,000 mg/L), 25 tests</a>

**Chloride**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Mercuric Thiocyanate, Method 8113	0.1 to 25.0 mg/L Cl <sup>-</sup>		<a href="#">View</a>		<a href="#">Water, Deionized 4 L</a>

**Chlorine, Free**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
DPD Rapid Liquid, Method 10059	0.02 to 2.00 mg/L		<a href="#">View</a>		
DPD, Method 10069	0.1 to 10.0 mg/L as Cl <sub>2</sub>		<a href="#">View</a>		<a href="#">DPD Free Chlorine Reagent, 25 mL Sample, Powder Pillows, PK/100</a>
DPD, Method 10102	0.09 to 5.00 mg/L		<a href="#">View</a>		<a href="#">DPD Free Chlorine Reagent, 10 mL Sample, Test 'N Tube Vials, PK/50</a>
DPD Method 8021	0.02 to 2.00 mg/L		<a href="#">View</a>	USEPA , USEPA , USEPA	<a href="#">DPD Free Chlorine Reagent, 10 mL Sample, Powder Pillows, PK/100</a> , <a href="#">DPD Free Chlorine Reagent, AccuVac® Ampules, PK/25</a>

**Chlorine, Total**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
DPD, Method 10014	2 to 500 µg/L as Cl <sub>2</sub>		<a href="#">View</a>	USEPA , USEPA , USEPA	<a href="#">Chlorine Reagent Set (Total), Ultra Low Range</a>
DPD Rapid Liquid, Method 10060	0.02 to 2.00 mg/L		<a href="#">View</a>		
DPD, Method 10070	0.1 to 10.0 mg/L as Cl <sub>2</sub>		<a href="#">View</a>	USEPA	<a href="#">DPD Total Chlorine Reagent, 25 mL Sample, Powder Pillows, PK/100</a>
DPD, Method 10101	0.09 to 5.00 mg/L		<a href="#">View</a>		<a href="#">DPD Total Chlorine Reagent, 10 mL Sample, Test 'N Tube Vials, PK/50</a>

**Environmental Bio-Systems, Inc. Field Laboratory  
Spectrophotometry Analysis DR-2800 Methods  
Parameter/Range/Reagent Information**

DPD, Method 8167	0.02 to 2.00 mg/L	<a href="#">View</a>	USEPA , USEPA , USEPA , USEPA , USEPA	<a href="#">DPD Total Chlorine Reagent, 10 mL Sample, Powder Pillows, PK/100 , DPD Total Chlorine Reagent, AccuVac® Ampules, PK/25 Chlorine Reagent Set (Total), Ultra Low Range , DPD Chlorine Indicator Solution, Ultra-Low Range, 1.5 mL Ampules, PK/20 , Buffer Solution, Ultra-Low Range Chlorine, pk/20, 1.5 mL Ampules , BLANKING REAGENT, 29ML</a>
DPD, Method 8370	2 to 500 µg/L as Cl <sub>2</sub>	<a href="#">View</a>	USEPA , USEPA , USEPA	

**Chlorine Dioxide**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Amaranth DPD, Method 10126	20 to 500 µg/L (Europe Only) 0.04 to 5.00 mg/L		<a href="#">View</a> <a href="#">View</a>	USEPA	<a href="#">Chlorine Dioxide DPD/Glycine AccuVac® Ampul Reagent Set, 25/test , Chlorine Dioxide Reagent Set, 10 mL Sample, 100/test Chlorine Dioxide Reagent 1 100 mL , Chlorine Dioxide Reagent 2 100 mL , Chlorine Dioxide Reagent 3 100 mL , Chlorine Dioxide Reagent Set, LR, 100/test Water, Deionized 4 L Water, Deionized 4 L</a>
Chlorophenol Red, Method 8065	0.01 to 1.00 mg/L		<a href="#">View</a>		
Direct Reading, Method 8138 Direct Reading, Method 8345	5 to 1000 mg/L 1 - 50 mg/L		<a href="#">View</a> <a href="#">View</a>		

**Chromium**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
1,5-Diphenylcarbohydrazide, Method 8023	0.010 to 0.700 mg/L Cr <sup>6+</sup>		<a href="#">View</a>	USEPA	<a href="#">Chromium Reagent Set, ChromaVer® , AccuVac® Ampuls, pk/25 , Chromium Reagent, ChromaVer® 3, Hexavalent, 5 mL or 10 mL sample, pk/100 Powder Pillow Chromium Total Reagent Set, 25 mL sample, 100/tests (Colorimetric Method) Chromium TNTplus (0.03-1.0 mg/L), 25 tests Chromium TNTplus (0.03-1.0 mg/L), 25 tests</a>
Alkaline Hypobromite Oxidation, Method 8024	0.01 to 0.70 mg/L		<a href="#">View</a>		
Diphenylcarbazide, Method 10218 Diphenylarbazide, Method 10219	0.03-1.00 mg/L Cr <sup>6+</sup> 0.03-1.00 mg/L Cr (Total)		<a href="#">View</a> <a href="#">View</a>	USEPA	

**Cobalt**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
1-2-Pyridylazo-2-Naphthol PAN, Method 8078	0.01 to 2.00 mg/L		<a href="#">View</a>		<a href="#">Cobalt / Nickel Reagent Set, PAN Method, 10- mL sample, 100 tests , EDTA REAGENT PWD PLWS PK/100 , PAN Indicator Solutions 100 mL , Phthalate-phosphate pp, pk/100</a>

**Color**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Platinum-Cobalt Standard, Method 8025	15 to 500 units		<a href="#">View</a>		<a href="#">Hydrochloric Acid Standard Solutions 1.0 N, 1.0 M 1 L , Sodium Hydroxide Standard Solutions 1.000 N 900 mL</a>

**Copper**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Porphyrim, Method 8143	1 to 210 µg/L		<a href="#">View</a>		<a href="#">Copper Reagent Set, 10 mL Sample, 100 tests , Porphyrim 1 Reagent Powder Pillows 50 tests , Porphyrim 2 Reagent Powder Pillows, 50 tests , Copper Masking Reagent Powder Pillows 100 tests</a>
Bicinchoninate, Method 8506	0.04 to 5.00 mg/L	0.04 mg/L	<a href="#">View</a>	USEPA , USEPA	<a href="#">Copper Reagent, CuVer® 1, 0 to 5.0 mg/L, 10 mL sample, pk/100 Powder Pillows Copper Reagent, CuVer® 2, 25/AccuVac® Ampules</a>
Bicinchoninate, Method 8026	0.04 to 5.00 mg/L	0.04 mg/L	<a href="#">View</a>		

**Cyanide**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
--------	-------	-----	-------	----------	------------------

**Environmental Bio-Systems, Inc. Field Laboratory  
Spectrophotometry Analysis DR-2800 Methods  
Parameter/Range/Reagent Information**

Pyridine-Pyrazalone, Method 8027	0.002 to 0.240 mg/L CN <sup>-</sup>	<a href="#">View</a>	<a href="#">Cyanide Reagent Sets, 100/tests</a> , <a href="#">CYANIVER 3 CYANIDE RGT PP PK/100</a> , <a href="#">CYANIVER 4 CYANIDE RGT PP PK/100</a> , <a href="#">CYANIVER 5 CYANIDE RGT PP PK/100</a>
----------------------------------	-------------------------------------	----------------------	-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

**Cyanuric Acid**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Turbidimetric, Method 8139	5 to 50 mg/L		<a href="#">View</a>		<a href="#">Cyanuric Acid 2 Reagent Powder Pillows 50</a>

**Dissolved Oxygen**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
HRDO, Method 8166	0.3 to 15.0 mg/L O <sub>2</sub>		<a href="#">View</a>		<a href="#">Dissolved Oxygen Reagent Set, High Range, AccuVac® Ampules, pk/25</a>
Indigo Carmine, Method 8316	6 to 800 µg/L O <sub>2</sub>		<a href="#">View</a>		<a href="#">Dissolved Oxygen Reagent Set, AccuVac® Ampules, Low Range, pk/25</a>
Ultra High Range, Method 8333	1.0 to 40.0 mg/L O <sub>2</sub>		<a href="#">View</a>		<a href="#">Dissolved Oxygen Reagent Set, High Range, AccuVac® Ampules, pk/25</a>

**Fluoride**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
SPADNS, Method 8029	0.02 to 2.00 mg/L F <sup>-</sup>		<a href="#">View</a>	USEPA , USEPA , USEPA , USEPA , USEPA , USEPA	<a href="#">SPADNS Fluoride Reagent AccuVac® Ampules, Pkg of 25</a> , <a href="#">SPADNS Fluoride Reagent Solution, 500 mL</a> , <a href="#">SPADNS 2 (Arsenic-free) Fluoride Reagent AccuVac® Ampules, Pkg of 25</a> , <a href="#">SPADNS 2 (Arsenic-free) Fluoride Reagent Solution, 500 mL</a> , <a href="#">SPADNS 2 (Arsenic-free) Fluoride Reagent Solution, 1 L</a> , <a href="#">SPADNS 2 (Arsenic-free) Fluoride Reagent Solution, 3.78 L</a>

**Formaldehyde**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
MBTH, Method 8110	3 to 500 µg/L		<a href="#">View</a>		<a href="#">REAGENT SET, FORMALDEHYDE LR</a>

**Hardness**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Calcium and Magnesium; Calmagite Colorimetric, Method 8030	0.05 to 4.00 mg/L Ca and Mg as CaCO <sub>3</sub>		<a href="#">View</a>		<a href="#">Hardness (Ca/Mg) Reagent Set, 100/tests</a>
Calcium and Magnesium; Chlorophosphonazo Colorimetric, Method 8374	8 to 1000 µg/L Ca and Mg as CaCO <sub>3</sub>		<a href="#">View</a>		<a href="#">Hardness (ULR) Reagent Set, Bulk Solution</a> , <a href="#">Hardness (ULR) Reagent Set, Solution Pillows, 100/PK</a>
Calcium and Magnesium; Chlorophosphonazo Colorimetric, Method 8374	4 to 1000 µg/L Ca & Mg as CaCO <sub>3</sub>		<a href="#">View</a>		<a href="#">CDTA Reagents, Bottle/10 mL</a> , <a href="#">Chlorophosphonazo Indicator Solution for ULR Hardness, 500mL</a>

**Hydrogen Sulfide**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Methylene Blue, Method 8131	5 to 800 µg/L		<a href="#">View</a>	USEPA	<a href="#">Hydrogen Sulfide Reagent Set, 0-0.600 mg/L as S<sub>2</sub>, 100 tests, Methylene Blue Method (Colorimetric) (Color Disc test: 0-2.25 mg/L)</a> , <a href="#">Sulfide 1 Reagent 100 tests MDB</a> , <a href="#">Sulfide 2 Reagent 100 tests MDB</a>

**Iodine**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
DPD Method 8031	0.07 to 7.00 mg/L		<a href="#">View</a>		<a href="#">DPD Total Chlorine Reagent, 10 mL Sample Powder Pillows, PK/100</a> , <a href="#">DPD Total Chlorine Reagent, AccuVac® Ampules, PK/25</a>

**Iron**

**Environmental Bio-Systems, Inc. Field Laboratory  
Spectrophotometry Analysis DR-2800 Methods  
Parameter/Range/Reagent Information**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
1, 10-Phenanthroline, Method 8146	0.02 to 3.00 mg/L		<a href="#">View</a>		<a href="#">Iron Ferrous Reagent, 1, 10 Phenanthroline, AccuVac® Ampuls, pk/25</a> , <a href="#">Iron Ferrous Reagent, 25ml Sample, pk/100 Powder Pillows</a>
FerroZine®, Method 8147	0.009 to 1.400 mg/L		<a href="#">View</a>		<a href="#">Iron Reagent Solution, FerroZine®, Solution Pillows, PK/50</a> , <a href="#">Iron Reagent, FerroZine®, (Rapid Liquid) Solution, 500 mL</a>
FerroZine® Rapid Liquid	0.009 to 1.400 mg/L Fe		<a href="#">View</a>		<a href="#">Iron Reagent, FerroZine®, (Rapid Liquid) Solution, 500 mL</a>
FerroVer® Rapid Liquid, Method 8147	0.02 to 3.00 mg/L		<a href="#">View</a>	USEPA	<a href="#">Iron (Total) Reagent, FerroVer®, 10mL Samples, PK/100</a> , <a href="#">Iron (Total) Reagent, FerroVer®, AccuVac®, pk/25</a>
TPTZ, Method 8112	0.012 to 1.800 mg/L		<a href="#">View</a>		<a href="#">Iron (Total) Reagent, (TPTZ) AccuVac® Ampules, pk/25</a> , <a href="#">Iron Reagent (TPTZ), 10mL Samples, pk/100</a>
FerroMo(TM), Method 8365	0.01 to 1.80 mg/L		<a href="#">View</a>		<a href="#">Iron (Total) Reagent Set, 0-1.80 mg/L, FerroMo Method, 100 tests</a> , <a href="#">Total Iron Reagent 1, pk/25 Powder Pillows (FerroMo Method)</a>

**Lead**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Dithizone, Method 8033	3 to 300 µg/L		<a href="#">View</a>	USEPA	<a href="#">Lead Reagent Sets 90 tests (No Air Shipments)</a>
LeadTrak Fast Column Extraction Method 8317	5 to 150 µg/L		<a href="#">View</a>		<a href="#">Lead Reagent Set, LeadTrak(TM) Fast Column Extraction Method, 20/tests</a>
PAR, Method 10216	0.1-2.0 mg/L Pb		<a href="#">View</a>		<a href="#">Lead TNTplus (0.10-2.0 mg/L), 25 tests</a>

**Manganese**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Periodate Oxidation, Method 8034	0.1 to 20.0 mg/L		<a href="#">View</a>	USEPA	<a href="#">Manganese Reagent Sets, HR, 100/tests</a> , <a href="#">Sodium Periodate Powder Pillows 100</a> , <a href="#">Buffer Solution for Manganese Citrate Type, cold periodate method, 10 mL sample, pk/100 Powder Pillows</a>
PAN, Method 8149	0.006 to 0.700 mg/L		<a href="#">View</a>		<a href="#">Ascorbic Acid, pk/100 Powder Pillows</a> , <a href="#">Manganese, Low Range, Reagent Set, 0-0.7 mg/L, 10 mL sample, 50/tests, PAN Method</a> , <a href="#">Alkaline Cyanide Reagent, 50 mL, Self-Contained Dropping Bottle</a> , <a href="#">PAN Indicator Solutions 50 mL, SCDB</a>

**Mercury**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Cold Vapor Mercury Preconcentration, Method 10065	0.1 to 2.5 µg/L		<a href="#">View</a>		<a href="#">Mercury Reagent Sets 25 tests</a>

**Molybdenum**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Mercaptoacetic Acid, Method 8036	0.2 to 40.0 mg/L		<a href="#">View</a>		<a href="#">HR Molybdenum Reagent Set, 10 mL Samples, 100 Tests</a>
Ternary Complex, Method 8169	0.02 to 3.00 mg/L		<a href="#">View</a>		<a href="#">LR Molybdenum Reagent Set, 20mL Samples, 100/tests</a>

**Monochloramine**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Indophenol, Method 10171	0.04 to 4.50 mg/L Cl <sub>2</sub>		<a href="#">View</a>		<a href="#">Monochlor F Reagent, pk/50 Powder Pillows</a>
Indopheno, Method 10172	0.1 to 10.0 mg/L Cl <sub>2</sub>		<a href="#">View</a>		<a href="#">Monochloramine Reagent Set, High Range (Indophenol Method), 0-10.0 mg/L, 50/tests</a>
Indophenol, Hach Method 10200	0.01-0.50 mg/L NH <sub>3</sub> -N		<a href="#">View</a>		<a href="#">Free Ammonia Reagent Set (Monochloramine/Free Ammonia, Indophenol Method), 50 tests</a>
	0.04-4.50 mg/L Cl <sub>2</sub>				

**Nickel**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
--------	-------	-----	-------	----------	------------------

ENVIRONMENTAL BIO-SYSTEMS ([www.ebsinfo.com](http://www.ebsinfo.com); [jimjacobs@ebsinfo.com](mailto:jimjacobs@ebsinfo.com))

**Environmental Bio-Systems, Inc. Field Laboratory  
Spectrophotometry Analysis DR-2800 Methods  
Parameter/Range/Reagent Information**

Heptoxime, Method 8037	0.02 to 1.80 mg/L Ni	<a href="#">View</a>	USEPA	<a href="#">Nickel Reagent Set, 0 - 1.80 mg/L as Ni, 50/test</a>
1-2 Pyridylazo-2-Naphthol PAN, Method 8150	0.006 to 1.000 mg/L	<a href="#">View</a>		<a href="#">Cobalt / Nickel Reagent Set, PAN Method, 10- mL sample, 100 tests, EDTA REAGENT PWD PLWS PK/100, PAN Indicator Solutions 100 mL, Phthalate-phosphate pp_pk/100</a>
Dimethylglyoxime, Method 10220	0.1-6.0 mg/L Ni	<a href="#">View</a>		<a href="#">Nickel TNTplus (0.1-6.0 mg/L), 25 tests</a>

**Nitrate**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Chromotropic Acid, Method 10020	0.2 to 30.0 mg/L NO <sub>3</sub> <sup>-</sup> -N		<a href="#">View</a>		<a href="#">Nitrogen-Nitrate Reagent Sets, 0-30 mg/L, 50/tests</a>
Cadmium Reduction, Method 8039	0.2 to 30.0 mg/L NO <sub>3</sub> <sup>-</sup> -N		<a href="#">View</a>		<a href="#">Nitrogen-Nitrate Reagent Sets, 0-30 mg/L, 50/tests</a>
Cadmium Reduction, Method 8171	0.1 to 10.0 mg/L NO <sub>3</sub> <sup>-</sup> -N		<a href="#">View</a>		<a href="#">Nitrate HR Reagent, NitraVer® 5, 10 mL sample, pk/100, Nitrate, AccuVac® Ampules, NitraVer® 5, pk/25</a>
Cadmium Reduction, Method 8192	0.01 to 0.50 mg/L NO <sub>3</sub> <sup>-</sup> -N		<a href="#">View</a>		<a href="#">Nitrogen, Nitrite, LR, NitriVer® 3 Reagent, 100/test, NitraVer® 6 Nitrate Reagent Powder Pillows, 10 mL sample, pk/100, Nitrogen-Nitrate Reagent Sets (LR), 10mL sample, 100 tests</a>
Dimethylphenol, Method 10206	0.23-13.50 mg/L NO <sub>3</sub> -N		<a href="#">View</a>		<a href="#">Nitrate TNTplus, LR (0.23-13.5 mg/L NO3-N), 25 tests</a>
Dimethylphenol, Method 10206	5-35 mg/L NO <sub>3</sub> -N		<a href="#">View</a>		<a href="#">Nitrate TNTplus, HR (5-35 mg/L NO3-N), 25 tests</a>

**Nitrite**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Diazotization, Method 10019	0.003 to 0.500 mg/L NO <sub>2</sub> <sup>-</sup> -N		<a href="#">View</a>		<a href="#">Nitrogen-Nitrite, Low Range, Reagent Set 50/tests</a>
Ferrous Sulfate, Method 8153	2 to 250 mg/L NO <sub>2</sub> <sup>-</sup>		<a href="#">View</a>		<a href="#">Nitrogen, Nitrite HR, NitriVer® 2 Reagent, 10 mL samples, pk/100 Powder Pillows</a>
Diazotization, Method 8507	0.002 to 0.300 mg/L NO <sub>2</sub> <sup>-</sup> -N		<a href="#">View</a>	USEPA	<a href="#">Nitrogen, Nitrite, LR, NitriVer® 3 Reagent, 100/test, Nitrogen, Nitrite, AccuVac® Ampules, NitriVer® 3, pk/25</a>
Diazotization, Method 10207	0.015-0.60 mg/L NO <sub>2</sub> -N		<a href="#">View</a>	USEPA	<a href="#">Nitrite TNTplus, Low Range (0.015-0.60 mg/L NO2-N), 25 tests</a>

**Nitrogen, Monochloramine and Free Ammonia**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
--------	-------	-----	-------	----------	------------------

**Nitrogen, Total**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Persulfate Digestion, Method 10071	0.5 to 25.0 mg/L N		<a href="#">View</a>		<a href="#">Nitrogen, Total Reagent Set, Test 'N Tube, 50/tests</a>
Persulfate Digestion, Method 10072	2 to 150 mg/L N		<a href="#">View</a>		<a href="#">Nitrogen, Total, High Range Reagent Set, 50/tests</a>
Total Nitrogen, Method 10208	1-16 mg/L N		<a href="#">View</a>		<a href="#">Nitrogen (Total) TNTplus, Low Range (1-16 mg/L), 25 tests</a>
Total Nitrogen, Method 10208	5-40 mg/L N		<a href="#">View</a>		<a href="#">Nitrogen (Total) TNTplus, High Range (5-40 mg/L), 25 tests</a>
Total Nitrogen, Method 10208	20-100 mg/L N		<a href="#">View</a>		<a href="#">Nitrogen (Total) TNTplus, Ultra High Range (20-100 mg/L), 25 tests</a>

**Nitrogen, Total Inorganic (TIN)**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Titanium Trichloride Reduction, Method 10021	0.2 to 25.0 mg/L N		<a href="#">View</a>		<a href="#">Nitrogen-Ammonia Reagent Set, Low-Range Ammonia, 50 ready to use tubes, Nitrogen-Inorganic, Total Pretreatment Reagent Set, 50/tests</a>

**Nitrogen, Total Kjeldahl (TKN)**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Nessler, Method 8075	1 to 150 mg/L		<a href="#">View</a>		<a href="#">Nitrogen (Total Kjeldahl) Reagent Set, 250 tests, TKN Indicator Solution 50 mL SCDB</a>

**Environmental Bio-Systems, Inc. Field Laboratory  
Spectrophotometry Analysis DR-2800 Methods  
Parameter/Range/Reagent Information**

**Nitrogen-Ammonia/Ammonium**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Salicylate Method, Method 10205	0.015-2.0 mg/L NH <sub>3</sub> -N		<a href="#">View</a>		Ammonia TNTplus, Ultra Low Range (0.015-2 mg/L NH <sub>3</sub> -N), 25 tests
Salicylate Method, Method 10205	1-12 mg/L NH <sub>3</sub> -N		<a href="#">View</a>		Ammonia TNTplus, Low Range (1-12 mg/L NH <sub>3</sub> -N), 25 tests
Salicylate Method, Method 10205	2-47 mg/L NH <sub>3</sub> -N		<a href="#">View</a>		Ammonia TNTplus, High Range (2-47 mg/L NH <sub>3</sub> -N), 25 tests
Salicylate, Method 10023	0.02 to 2.50 mg/L NH <sub>3</sub> -N		<a href="#">View</a>		Nitrogen-Ammonia Reagent Set, Low-Range Ammonia, 50 ready to use tubes
Salicylate, Method 10031	0.4 to 50.0 mg/L NH <sub>3</sub> -N		<a href="#">View</a>		Nitrogen-Ammonia Reagent Set, HR, Test 'N Tube, Salicylate Method, 0.4-50 mg/L, 50/tests
Nessler, Method 8038	0.02 to 2.50 mg/L NH <sub>3</sub> -N		<a href="#">View</a>		Mineral Stabilizer Solution 50 mL SCDB, Nessler Reagent 500 mL, Nitrogen-Ammonia Reagent Set, 0-3 mg/L, Nessler's Method, Colorimetric, 250/tests, Polyvinyl Alcohol Dispersing Agent 50 mL SCDB
Salicylate, Method 8155	0.01 to 0.50 mg/L NH <sub>3</sub> -N		<a href="#">View</a>		Ammonia Cyanurate Reagent, pk/100 Powder Pillows, Ammonia Salicylate Reagent, 10 mL sample, pk/100 Powder Pillows, Nitrogen-Ammonia Reagent Set, 0-0.5 mg/L as NH <sub>3</sub> -N, 10 mL sample, 100 tests
Indophenol, Method 10201	0.01 to 0.50 mg/L NH <sub>3</sub> -N		<a href="#">View</a>		Free Ammonia Reagent Set (Monochloramine/Free Ammonia, Indophenol Method), 50 tests, Free Ammonia Reagent Solution, 4ml SCDB, Monochlor F Reagent, pk/100 powder pillows

**Oxygen Scavengers**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Iron Reduction, Method 8140	5 - 600 µg/L carbohydrazide		<a href="#">View</a>		Oxygen Scavenger Reagent Set, 25 mL samples, 100/test
	3 - 450 µg/L DEHA				
	9 -1000 µg/L hydroquinone				
	13 -1500 µg/L ISA				
p-Dimethylaminobenzaldehyde, Method 8141	15 - 1000 µg/L MEKO 4 to 600 µg/L		<a href="#">View</a>		Hydrazine Reagent, AccuVac® Ampuls, pk/25, Hydrazine Reagent, HydraVer® 2, 100 mL

**Ozone**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Indigo, Method 8311	0.01 to 0.25 mg/L O <sub>3</sub>		<a href="#">View</a>		Ozone Reagent, LR, AccuVac® Ampuls, 0-0.25 mg/L, pk/25
Indigo, Method 8311	0.01 to 0.75 mg/L O <sub>3</sub>		<a href="#">View</a>		Ozone Reagent, MR, AccuVac® Ampuls, pk/25
Indigo, Method 8311	0.01 to 1.50 mg/L O <sub>3</sub>		<a href="#">View</a>		Ozone Reagent, HR, AccuVac® Ampuls, pk/25

**Phenols**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
4-Aminoantipyrine, Method 8047	0.002 to 0.200 mg/L		<a href="#">View</a>	USEPA	Chloroform 4 L (No Air Shipments), Hardness1 Buffer Solution, 500 mL, (pH10.1), Phenol 2 Reagent, pk/100 Powder Pillows, Phenol Reagent, pk/100 Powder Pillows, Phenols Reagent Set 100 tests (No Air Shipments)

**Phosphonates**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
--------	-------	-----	-------	----------	------------------

**Environmental Bio-Systems, Inc. Field Laboratory  
Spectrophotometry Analysis DR-2800 Methods  
Parameter/Range/Reagent Information**

Persulfate UV Oxidation, Method 8007	0.02 to 2.50 and 1.0	<a href="#">View</a>		<a href="#">Phosphate, PhosVer® 3 Reagent, 10 mL sample, pk/100 Powder Pillows, Phosphonates Reagent Set, 100/tests, Potassium Persulfate Powder Pillows 100 Phosphorus TNTplus, LR (0.05-1.50 mg/L PO4-P), 25 tests</a>
Phosphomolybdate, Method 10209	0.05-1.5 mg/L PO <sub>4</sub> -P	<a href="#">View</a>		
Phosphomolybdate, Method 10209	0.15-4.50 mg/L PO <sub>4</sub> 0.5-5.0 mg/L PO <sub>4</sub> -P	<a href="#">View</a>		<a href="#">Phosphorus TNTplus, HR (0.5-5.0 mg/L PO4-P), 25 tests</a>
Phosphomolybdate, Method 10209	1.5-15.0 mg/L PO <sub>4</sub> 2-20 mg/L PO <sub>4</sub> -P	<a href="#">View</a>		<a href="#">Phosphorus TNTplus, UHR (2-20 mg/L PO4-P), 25 tests</a>
Phosphomolybdate, Method 10210	6-60 mg/L PO <sub>4</sub> 0.05-1.50 mg/L PO <sub>4</sub> -P	<a href="#">View</a>	USEPA	<a href="#">Phosphorus TNTplus, LR (0.05-1.50 mg/L PO4-P), 25 tests</a>
Phosphomolybdate, Method 10210	0.15-4.50 mg/L PO <sub>4</sub> 2-20 mg/L PO <sub>4</sub> -P	<a href="#">View</a>	USEPA	<a href="#">Phosphorus TNTplus, UHR (2-20 mg/L PO4-P), 25 tests</a>
Molybdoanadate, Method 10214	6-60 mg/L PO <sub>4</sub> 1.6-30.0 mg/L PO <sub>4</sub> -P	<a href="#">View</a>		<a href="#">Phosphorus, Reactive (Ortho) TNTplus (1.67-30.0 mg/L PO4-P), 25 tests</a>
Phosphomolybdate, Method 10210	0.5-5.0 mg/L PO <sub>4</sub> -P	<a href="#">View</a>	USEPA	<a href="#">Phosphorus TNTplus, HR (0.5-5.0 mg/L PO4-P), 25 tests</a>
	1.5-15.0 mg/L PO <sub>4</sub>			

**Phosphorus**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Acid Digestion, Method 8180	0.06 to 3.50 mg/L PO <sub>4</sub> <sup>3-</sup>		<a href="#">View</a>	USEPA	<a href="#">Phosphate (Acid Hydrolyzable) Reagent Set, 50/tests, Phosphate Reagent, PhosVer® 3, 10 mL sample, pk/50 Powder Pillows, POTASSIUM PERSULFATE PP PK/50, SODIUM HYDROXIDE, 1.54N 100ML Sodium Hydroxide Standard Solutions 5.0 N 100 tests MDB, Sulfuric Acid Standard Solutions, 5.25 N, 100 tests MDB</a>
PhosVer® 3 with Acid Hydrolysis, Method 8180	Acid Digestion		<a href="#">View</a>	USEPA	<a href="#">Phosphorus Rapid Liquid Reagent Set, Ascorbic Acid, 250/test</a>
Ascorbic Acid Rapid Liquid, Method 10055	19 to 3000 µg/L PO <sub>4</sub> <sup>3-</sup>		<a href="#">View</a>		<a href="#">Phosphate Reagent, PhosVer® 3, 10 mL sample, pk/50 Powder Pillows, Phosphorus, Reactive (Orthophosphate), UR, 50/tests</a>
PhosVer® 3 Ascorbic Acid, Method 8048	0.06 to 5.00 mg/L PO <sub>4</sub> <sup>3-</sup> or		<a href="#">View</a>	USEPA, USEPA, USEPA	<a href="#">Phosphate, PhosVer® 3 Reagent, 10 mL sample, pk/100 Powder Pillows, Phosphate Reagent, PhosVer® 3, pk/25 AccuVac® Ampules</a>
PhosVer® 3 Ascorbic Acid, Method 8048	0.02 to 1.60 mg/L P 0.02 to 2.50 mg/L PO <sub>4</sub> <sup>3-</sup>		<a href="#">View</a>	USEPA, USEPA, USEPA	<a href="#">Phosphate, PhosVer® 3 Reagent, 10 mL sample, pk/100 Powder Pillows, Phosphate Reagent, PhosVer® 3, pk/25 AccuVac® Ampules</a>
PhosVer® 3, Method 8048	0.02 to 2.50 mg/L PO <sub>4</sub> <sup>3-</sup>		<a href="#">View</a>	USEPA, USEPA, USEPA	<a href="#">Phosphate Reagent, PhosVer® 3, 10 mL sample, pk/50 Powder Pillows, Phosphorus, Reactive (Orthophosphate), UR, 50/tests</a>
PhosVer® 3, Method 8048	0.06 to 5.00 mg/L PO <sub>4</sub> <sup>3-</sup> or 0.02 to 1.60 mg/L P		<a href="#">View</a>	USEPA, USEPA, USEPA	<a href="#">Phosphorus, Rapid Liquid(TM) Reagents 500 mL, Phosphorus, Test 'N Tube, HR, Reative, Reagent Set, 50/test, Molybdoanadate Reagent, bottle/100 mL MDB, Phosphorus, Molybdoanadate, AccuVac® Ampules, pk/25, Amino Acid Reagent 100 tests, Molybdate Reagent 100 mL MDB, Phosphorus, Reactive (Orthophosphate) Reagent Set, 100/tests, Sodium Hydroxide Standard Solutions 5.0 N 100 tests MDB, Sulfuric Acid Standard Solutions, 5.25 N, 100 tests MDB, Potassium Persulfate Powder Pillows 100</a>
Molybdoanadate, Method 8114	0.3 to 45.0 mg/L PO <sub>4</sub> <sup>3-</sup>		<a href="#">View</a>		<a href="#">POTASSIUM PERSULFATE PP PK/50, SODIUM HYDROXIDE, 1.54N 100ML, MOLYBDOVANADATE RGT, 25ML,</a>
Molybdoanadate, Method 8114	1.0 to 100.0 mg/L PO <sub>4</sub> <sup>3-</sup>		<a href="#">View</a>		
Molybdoanadate Rapid Liquid, Method 8114	0.3 to 45.0 mg/L PO <sub>4</sub> <sup>3-</sup>		<a href="#">View</a>		
Amino Acid, Method 8178	0.23 to 30.00 mg/L PO <sub>4</sub> <sup>3-</sup>		<a href="#">View</a>		
Acid Persulfate Digestion, Method 8190	Acid Persulfate Digestion		<a href="#">View</a>	USEPA	
Molybdoanadate, Method 10127	1.0 to 100.0 mg/L PO <sub>4</sub> <sup>3-</sup>		<a href="#">View</a>		

**Environmental Bio-Systems, Inc. Field Laboratory  
Spectrophotometry Analysis DR-2800 Methods  
Parameter/Range/Reagent Information**

PhosVer® 3 with Acid Persulfate Digestion, Method 8190	0.06 to 3.50 mg/L PO <sub>4</sub> <sup>3-</sup> - or  0.02 to 1.10 mg/L P	<a href="#">View</a>	USEPA	<a href="#">Phosphorus, Total High Range Test 'N Tube Reagent Set, 50/test</a> <a href="#">Phosphate Reagent, PhosVer® 3, 10 mL sample, pk/50 Powder Pillows,</a> <a href="#">POTASSIUM PERSULFATE PP PK/50,</a> <a href="#">SODIUM HYDROXIDE, 1.54N 100ML,</a> <a href="#">Phosphorus, Total, Test 'N Tube Reagent Set, 50/tests</a>
--------------------------------------------------------	---------------------------------------------------------------------------------	----------------------	-------	---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

**Polychlorinated Bi-Phenols (PCB)**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Immunoassay, Method 10050	Immunoassay Method		<a href="#">View</a>		

**Potassium**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Tetraphenylborate, Method 8049	0.1 to 7.0 mg/L		<a href="#">View</a>		<a href="#">Potassium 1 Reagent, pk/25 Powder Pillows,</a> <a href="#">Potassium 2 Reagent, pk/25 Solution Pillows,</a> <a href="#">Potassium 3 Reagent, pk/100 Powder Pillows,</a> <a href="#">REAGENT SET, POTASSIUM 0.25-2.50 mg/L as K</a>

**Quaternary Ammonium Compounds (QAC)**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Direct Binary Complex, Method 8337	0.2 to 5.0 mg/L as CTAB		<a href="#">View</a>		<a href="#">QAC REAGENT 1 PWD PLW PK/50,</a> <a href="#">QAC REAGENT 2, PWD PLWS PK/25</a>

**Selenium**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Diaminobenzidine, Method 8194	0.01 to 1.00 mg/L		<a href="#">View</a>		<a href="#">Buffer Solution, Sulfate Type 500 mL,</a> <a href="#">Diaminobenzidine, Tetrahydrochloride 5 g,</a> <a href="#">Potassium Hydroxide Standard Solutions 12, 0 N 100 tests MDB,</a> <a href="#">REAGENT SET, FOR SELENIUM,</a> <a href="#">TitraVer® Hardness Reagent 100 g,</a> <a href="#">Toluene 4 L (No Air Shipments)</a>

**Silica**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Silicomolybdate, Method 8185	1 to 100 mg/L		<a href="#">View</a>		<a href="#">Acid Reagent Powder Pillows for HR Silica, 10mL sample, pk/100,</a> <a href="#">Citric Acid for Silica, 10 mL Sample, pk/100 Powder Pillows,</a> <a href="#">Molybdate Reagent Powder Pillows for HR Silica, 10 mL sample, pk/100,</a> <a href="#">Silica Reagent Set, High Range, 10 mL Samples, 100/tests</a>
Heteropoly Blue, Method 8186	0.010 to 1.600 mg/L as SiO <sub>2</sub>		<a href="#">View</a>		<a href="#">Citric Acid for Silica, 10 mL Sample, pk/100 Powder Pillows,</a> <a href="#">Amino Acid F Reagent Powder Pillows, 10 mL sample, pk/100 (for silica analysis),</a> <a href="#">Molybdate 3 Reagent Solution, 50 mL, SCDB,</a> <a href="#">Silica, Low Range, Reagent Set, 10mL Samples, 100 Tests</a>
Heteropoly Blue, Method 8282	3 to 1000 µg/L as SiO <sub>2</sub>		<a href="#">View</a>		<a href="#">Amino Acid F Reagent Solution, 100mL,</a> <a href="#">Amino Acid F Reagent Solution, 20 Ampules (for ultra low range silica analysis),</a> <a href="#">Citric Acid F Reagent Solution, for Rapid Liquid Silica, 500 mL,</a> <a href="#">Silica, Ultra Low Range, Bulk Solution, Reagent Set, 100 tests,</a> <a href="#">Silica, Ultra Low Range, Reagent Set, 40/test</a>
Heteropoly Blue, Method 8282	3 to 1000 µg/L as SiO <sub>2</sub>		<a href="#">View</a>		<a href="#">Citric Acid F Reagent Solution, for Rapid Liquid Silica, 500 mL,</a> <a href="#">Amino Acid F Dilution Solvent, for Rapid Liquid Silica, 475 mL,</a> <a href="#">Amino Acid F Reagent Powder, for Rapid Liquid Silica, 55 g,</a> <a href="#">Molybdate 3 Reagent, for Rapid Liquid Silica, 500 mL</a>

**Environmental Bio-Systems, Inc. Field Laboratory  
Spectrophotometry Analysis DR-2800 Methods  
Parameter/Range/Reagent Information**

**Silver**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Colorimetric, Method 8120	0.005 to 0.700 mg/L		<a href="#">View</a>		REAGENT SET, SILVER COLORIMETRIC METHOD, Silver 1 Reagent Powder Pillows pk/50 , Silver 2 Reagent Solution Pillows pk/50 , Sodium Thiosulfate Powder Pillows 50/pk

**Sulfate**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
SulfaVer® 4, Method 8051	2 to 70 mg/L		<a href="#">View</a>	, USEPA	Sulfate Reagent, SulfaVer® 4, 10-mL sample_pk/100 Powder Pillows , Sulfate Reagent, SulfaVer® 4, pk/25 AccuVac® Ampules

**Surfactants, Anionic**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Crystal Violet, Method 8028	0.002 to 0.275 mg/L as LAS		<a href="#">View</a>		Buffer Solution, Sulfate Type 500 mL , BENZENE, ACS 4L , Detergents Reagent Powder Pillows 25 , Detergents Reagent Sets (No Air Shipments)

**Suspended Solids**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Photometric, Method 8006	5 to 750 mg/L		<a href="#">View</a>		

**Tannin and Lignin**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Tyrosine, Method 8193	0.1 to 9.0 mg/L		<a href="#">View</a>		Sodium Carbonate Solution 500 mL , Tannin-Lignin Reagent Sets, up to 100 tests , TanniVer® 3 Tannin-Lignin Reagent 100 tests MDB

**Total Organic Carbon (TOC)**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Direct, Method 10128	100 to 700 mg/L C		<a href="#">View</a>		Total Organic Carbon (TOC) Reagent Set, High Range, 100-700 mg/L, 50/test
Direct, Method 10129	0.3 to 20.0 mg/L C		<a href="#">View</a>		Total Organic Carbon (TOC) Reagent Set, Low Range, 50/test
Direct, Method 10173	15 to 150 mg/L C		<a href="#">View</a>		Total Organic Carbon (TOC) Reagent Set, Mid Range, 50/test

**Total Petroleum Hydrocarbons (TPH)**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Immunoassay, Method 10050	Immunoassay Method		<a href="#">View</a>		

**Toxicity**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
ToxTrak(TM), Method 10017	0 to 100% Inhibition		<a href="#">View</a>		Reagent Set, ToxTrak Toxicity, 25-49 tests

**Triazole**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
UV Photolysis, Method 8079	Benzotriazole 1.0 to 16.0 mg/L  Tolyltriazole 1.0 to 20.0 mg/L		<a href="#">View</a>		Triazole Reagent, pk/100 Powder Pillows

**Trihalomethanes, Total (THM)**

**Environmental Bio-Systems, Inc. Field Laboratory  
Spectrophotometry Analysis DR-2800 Methods  
Parameter/Range/Reagent Information**

---

Method	Range	EDL	Proc.	Approved	Reagent Ordering
THM Water Bath, Method 10132	10 to 600 ppb as Chloroform		<a href="#">View</a>		<a href="#">THM Reagent 1 Solution, 30 mL, SCDB</a> , <a href="#">THM Reagent 2 Solution, 330 mL</a> , <a href="#">THM Reagent 3 Solution, 100 mL</a> , <a href="#">THM Reagent 4, pk/100 Powder Pillows</a> , <a href="#">Total Trihalomethanes Reagent Set</a> (Colorimetric method) (THM) 100 tests

**Volatile Acids**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Esterification, Method 8196	27 to 2800 mg/L		<a href="#">View</a>		<a href="#">Volatile Acids Reagent Set, 90 tests</a>

**Zinc**

Method	Range	EDL	Proc.	Approved	Reagent Ordering
Zincon, Method 8009	0.01 to 3.00 mg/L		<a href="#">View</a>	USEPA ,	<a href="#">Zinc Reagent Set, 20-mL sample size, 100 tests</a>